



EC - TYPE EXAMINATION CERTIFICATE

SK 08 - 001 MI-010

Rev. 0 Add. 5

Issued by

Slovenská legálna metrologia, n. o.
Hviezdoslavova 31
974 01 Banská Bystrica
Slovak Republic

Notified body **1432**

In accordance with

Government Ordinance of the Slovak Republic No 294/2005 Coll. on measuring instruments as amended by Government Ordinance No 445/2010 Coll., which implements, in the Slovakia, the Directive 2004/22/EC of the European Parliament and of the Council of 31 March 2004 on measuring instruments as later amended (MID).

Applicable essential requirements

Annex I and Annex MI-010 to MID

Manufacturer

ACTIA CZ s.r.o.
Lesní 47
390 01 Tábor - Horky, Česká republika

Applicant

Manufacturer

Measuring instrument

exhaust gas analyser

Type

AT 505

Trade mark

see Descriptive Annex

SW version

1.1.x or 1.2.x

Environment classes

- climatic

+5 to +40 °C

- mechanical

M2

- electromagnetic

E2

Description and documentation

The principal technical and metrological data, characteristics, instrument description and approval conditions are set out in the Final Protocol No. 66/1432/14 MI-010 to this EC - type examination certificate (2 pages), which is part of this EC - type examination certificate. The test reports, designs, schematic diagrams and documentation used during certification process are recorded under reference folder Atal_AT505_00_05.

Valid until

8 February 2018

Date of issue

31 December 2014



Reg. No. 058/P-017



Ing. Stefan Král, PhD.

Representative of Notified Body

Where the instrument is subject to other Directives covering other aspects, this EC - type examination certificate is valid, assuming that the instrument conforms to the provisions of those Directives. Without written permission of the notified body this certificate may be reproduced only as a whole.

